

Supplement

Manual Title:	80PK-11 Instruction Sheet		
Part Number:	2724595	Supplement Issue:	1
Print Date:	August 2006	Issue Date:	6/08
Revision/Date:		Page Count:	1

This supplement contains
information necessary to ensure
the accuracy of the above
manual.

Change #1, 47083

Under ***Specifications***, change the following:

From: **Measurement Range:** -30 °C to 105 °C
(-22 °F to 221 °F)

To: **Measurement Range:** -30 °C to 150 °C
(-19.3 °F to 302 °F)

From: **Display Accuracy** ±0.1 °C (0.1 °F)

To: **Accuracy** ±1.5 °C (2.7 °F)

Add the following:

Operating Temperature: 0 °C to 50 °C

Storage Temperature: -20 °C to 65 °C